



**SISCO
RESEARCH
LABORATORIES
PVT. LTD.**

www.srlchem.com

Certificate of Analysis

Product

74399 - Buffer capsules pH 9.2 (10 capsules)

Batch No

3607230

Analysis Date

21-May-2024

Date of Manufacture

May 2024

Expiry Date/Re-Test Date

May 2029

Test Parameters	Standards	Actual Results
pH	9.2 +/- 0.05 (Dissolve contents of 1 capsule & dilute upto 100 ml distilled water to prepare solution of pH 9.2 at 20°C)	9.19

FRM/82/11/4/74399 Rev No. 1 Date 07/18/2017

Prepared by **Prajakta**

In compliance with the standard specifications of SRL

Plant Site 2 : H-4, MIDC, Talaja – 410 208, Maharashtra, India

608-B, Satellite Gazebo, Andheri Ghatkopar Link Road,
Chakala, Andheri (E), Mumbai - 400 099. India. ☎ +91-22-4268 5800

1/1

[an ISO 9001-2015, 17025:2017, CE, GMP Certified Company](#)

This COA has been generated electronically and it is valid without signature.